METHOD #: 208.2	Approved for NPDES and SDWA (Issued 1978)
TITLE:	Barium (AA, Furnace Technique)
ANALYTE:	CAS # Ba Barium 7440-39-3
INSTRUMENTATION:	AA
STORET No.	01007 Dissolved 01005 Suspended 01006
Optimum Concentration Range: Detection Limit:	10-200 μg/L 2 μg/L

- 1.0 Preparation of Standard Solution
 - 1.1 Stock solution: Prepare as described under "direct aspiration method".
 - 1.2 Prepare dilutions of the stock solution to be used as calibration standards at the time of analysis. These solutions are also to used for "standard additions".
 - 1.3 The calibration standard should be diluted to contain 0.5% (v/v) HNO₃.
- 2.0 Sample Preservation
 - 2.1 For sample handling and preservation, see part 4.1 of the Atomic Absorption Methods section of this manual.
- 3.0 Sample Preparation
 - 3.1 Prepare as described under "direct aspiration method". Sample solutions for analysis should contain 0.5% (v/v) HNO₃.
- 4.0 Instrument Parameters (General)
 - 4.1 Drying Time and Temp: 30 sec-125°C.
 - 4.2 shing Time and Temp: 30 sec-1200°C.
 - 4.3 Atomizing Time and Temp: 10 sec-2800°C.
 - 4.4 Purge Gas Atmosphere: Argon
 - 4.5 Wavelength: 553.6 nm
 - 4.6 Other operating parameters should be set as specified by the particular instrument manufacturer.
- 5.0 Analysis Procedure
 - 5.1 For the analysis procedure and the calculation, see "Furnace Procedure" part 9.3 of the Atomic Absorption Methods section of this manual.
- 6.0 Notes

- 6.1 The above concentration values and instrument conditions are for a Perkin-Elmer HGA-2100, based on the use of a 20 μ L injection continuous flow purge gas and pyrolytic graphite.
- 6.2 The use of halide acid should be avoided.
- 6.3 Because of possible chemical interaction, nitrogen should not be used as a purge gas.
- 6.4 For every sample matrix analyzed, verification is necessary to determine that method of standard addition is not required (see part 5.2.1 of the Atomic Absorption Methods section of this manual).
- 6.5 If method of standard addition is required, follow the procedure given earlier in part 8.5 of the Atomic Absorption Methods section of this manual.
- 6.6 For quality control requirements and optional recommendations for use in drinking water analyses, see part 10 of the Atomic Absorption Methods section of this manual.
- 6.7 Data to be entered into STORET must be reported as $\mu g/L$.
- 7.0 Precision and Accuracy
 - 7.1 In a single laboratory (EMSL), using Cincinnati, Ohio tap water spiked at concentrations of 500 and 1000 μ g Ba/L, the standard deviations were ± 2.5 and ± 2.2 μ g, respectively. Recoveries at these levels were 96% and 102%, respectively A dilution of 1:10 was required to bring the spikes within the analytical range of the method.